

Continuing Calibration Report for SequenceID : FC073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC066859.D

Aliphatic C9-C12	7123629.000	60.000	3.184	6.469	118727.150	146546.385	18.983
Aliphatic C12-C16	4925742.000	40.000	6.470	9.865	123143.550	147989.695	16.789
Aliphatic C16-C21	7257121.000	60.000	9.866	13.227	120952.017	141330.589	14.419
Aliphatic C21-C28	9405873.000	80.000	13.228	16.885	117573.413	130900.885	10.181
Aliphatic C28-C40	14679964.000	120.000	16.886	21.755	122333.033	119057.916	-2.751
Aliphatic EPH	43392329.000	360.000	3.184	21.755	120534.247	133197.853	9.507