

Continuing Calibration Report for SequenceID : FC073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC066866.D

Aliphatic C9-C12	9361356.000	60.000	3.184	6.469	156022.600	146546.385	-6.466
Aliphatic C12-C16	6239286.000	40.000	6.470	9.865	155982.150	147989.695	-5.401
Aliphatic C16-C21	9027729.000	60.000	9.866	13.227	150462.150	141330.589	-6.461
Aliphatic C21-C28	11199919.000	80.000	13.228	16.885	139998.988	130900.885	-6.950
Aliphatic C28-C40	14884862.000	120.000	16.886	21.755	124040.517	119057.916	-4.185
Aliphatic EPH	50713152.000	360.000	3.184	21.755	140869.867	133197.853	-5.760