

Continuing Calibration Report for SequenceID : FC073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC066871.D

Aliphatic C9-C12	8329650.000	60.000	3.184	6.469	138827.500	146546.385	5.267
Aliphatic C12-C16	5469025.000	40.000	6.470	9.865	136725.625	147989.695	7.611
Aliphatic C16-C21	7831381.000	60.000	9.866	13.227	130523.017	141330.589	7.647
Aliphatic C21-C28	9744189.000	80.000	13.228	16.885	121802.363	130900.885	6.951
Aliphatic C28-C40	13385518.000	120.000	16.886	21.755	111545.983	119057.916	6.309
Aliphatic EPH	44759763.000	360.000	3.184	21.755	124332.675	133197.853	6.656