

Continuing Calibration Report for SequenceID : FC073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC066877.D

Aliphatic C9-C12	8397099.000	60.000	3.184	6.469	139951.650	146546.385	4.500
Aliphatic C12-C16	5487427.000	40.000	6.470	9.865	137185.675	147989.695	7.301
Aliphatic C16-C21	7867887.000	60.000	9.866	13.227	131131.450	141330.589	7.217
Aliphatic C21-C28	9815506.000	80.000	13.228	16.885	122693.825	130900.885	6.270
Aliphatic C28-C40	13496994.000	120.000	16.886	21.755	112474.950	119057.916	5.529
Aliphatic EPH	45064913.000	360.000	3.184	21.755	125180.314	133197.853	6.019