

Continuing Calibration Report for SequenceID : FD073124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048090.D

Aromatic C10-C12	6900942.000	40.000	4.124	5.845	172523.550	170133.608	-1.405
Aromatic C12-C16	10456029.000	60.000	5.846	8.453	174267.150	172773.386	-0.865
Aromatic C16-C21	12546462.000	80.000	8.454	12.718	156830.775	159315.524	1.560
Aromatic C21-C36	27153393.000	180.000	12.719	18.123	150852.183	144053.653	-4.719
Aromatic EPH	57056826.000	360.000	4.124	18.123	158491.183	155129.575	-2.167