

Continuing Calibration Report for SequenceID : FD073124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048100.D

Aromatic C10-C12	6715650.000	40.000	4.124	5.845	167891.250	170133.608	1.318
Aromatic C12-C16	10161919.000	60.000	5.846	8.453	169365.317	172773.386	1.973
Aromatic C16-C21	12330296.000	80.000	8.454	12.718	154128.700	159315.524	3.256
Aromatic C21-C36	27203734.000	180.000	12.719	18.123	151131.856	144053.653	-4.914
Aromatic EPH	56411599.000	360.000	4.124	18.123	156698.886	155129.575	-1.012