

Continuing Calibration Report for SequenceID : FD073124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048106.D

Aromatic C10-C12	6973987.000	40.000	4.124	5.845	174349.675	170133.608	-2.478
Aromatic C12-C16	10477270.000	60.000	5.846	8.453	174621.167	172773.386	-1.069
Aromatic C16-C21	12606852.000	80.000	8.454	12.718	157585.650	159315.524	1.086
Aromatic C21-C36	27755940.000	180.000	12.719	18.123	154199.667	144053.653	-7.043
Aromatic EPH	57814049.000	360.000	4.124	18.123	160594.581	155129.575	-3.523