

Continuing Calibration Report for SequenceID : FE073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE049322.D

Aliphatic C9-C12	8176480.000	60.000	3.193	6.811	136274.667	152940.213	10.897
Aliphatic C12-C16	5559039.000	40.000	6.812	10.246	138975.975	151715.035	8.397
Aliphatic C16-C21	8208908.000	60.000	10.247	13.609	136815.133	146082.286	6.344
Aliphatic C21-C28	10915136.000	80.000	13.610	17.267	136439.200	140208.524	2.688
Aliphatic C28-C40	19383749.000	120.000	17.268	22.144	161531.242	147288.840	-9.670
Aliphatic EPH	52243312.000	360.000	3.193	22.144	145120.311	146948.039	1.244