

Continuing Calibration Report for SequenceID : FE073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE049328.D

Aliphatic C9-C12	8065868.000	60.000	3.193	6.811	134431.133	152940.213	12.102
Aliphatic C12-C16	5483874.000	40.000	6.812	10.246	137096.850	151715.035	9.635
Aliphatic C16-C21	8046618.000	60.000	10.247	13.609	134110.300	146082.286	8.195
Aliphatic C21-C28	10745543.000	80.000	13.610	17.267	134319.288	140208.524	4.200
Aliphatic C28-C40	18959059.000	120.000	17.268	22.144	157992.158	147288.840	-7.267
Aliphatic EPH	51300962.000	360.000	3.193	22.144	142502.672	146948.039	3.025