

Continuing Calibration Report for SequenceID : FE073124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE049345.D

Aliphatic C9-C12	9564140.000	60.000	3.193	6.811	159402.333	152940.213	-4.225
Aliphatic C12-C16	6364763.000	40.000	6.812	10.246	159119.075	151715.035	-4.880
Aliphatic C16-C21	9230902.000	60.000	10.247	13.609	153848.367	146082.286	-5.316
Aliphatic C21-C28	11989749.000	80.000	13.610	17.267	149871.863	140208.524	-6.892
Aliphatic C28-C40	19292148.000	120.000	17.268	22.144	160767.900	147288.840	-9.151
Aliphatic EPH	56441702.000	360.000	3.193	22.144	156782.506	146948.039	-6.692