

Continuing Calibration Report for SequenceID : FE080124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE049348.D

Aliphatic C9-C12	8657345.000	60.000	3.191	6.807	144289.083	152940.213	5.657
Aliphatic C12-C16	5900008.000	40.000	6.808	10.242	147500.200	151715.035	2.778
Aliphatic C16-C21	8632013.000	60.000	10.243	13.605	143866.883	146082.286	1.517
Aliphatic C21-C28	11565938.000	80.000	13.606	17.261	144574.225	140208.524	-3.114
Aliphatic C28-C40	20329888.000	120.000	17.262	22.136	169415.733	147288.840	-15.023
Aliphatic EPH	55085192.000	360.000	3.191	22.136	153014.422	146948.039	-4.128