

Continuing Calibration Report for SequenceID : FE080124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE049359.D

Aliphatic C9-C12	7349513.000	60.000	3.191	6.807	122491.883	152940.213	19.909
Aliphatic C12-C16	5050765.000	40.000	6.808	10.242	126269.125	151715.035	16.772
Aliphatic C16-C21	7417717.000	60.000	10.243	13.605	123628.617	146082.286	15.371
Aliphatic C21-C28	9913222.000	80.000	13.606	17.261	123915.275	140208.524	11.621
Aliphatic C28-C40	17482598.000	120.000	17.262	22.136	145688.317	147288.840	1.087
Aliphatic EPH	47213815.000	360.000	3.191	22.136	131149.486	146948.039	10.751