

Continuing Calibration Report for SequenceID : FE080124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE049375.D

Aliphatic C9-C12	8160185.000	60.000	3.191	6.807	136003.083	152940.213	11.074
Aliphatic C12-C16	5754572.000	40.000	6.808	10.242	143864.300	151715.035	5.175
Aliphatic C16-C21	8535227.000	60.000	10.243	13.605	142253.783	146082.286	2.621
Aliphatic C21-C28	11555150.000	80.000	13.606	17.261	144439.375	140208.524	-3.018
Aliphatic C28-C40	20840342.000	120.000	17.262	22.136	173669.517	147288.840	-17.911
Aliphatic EPH	54845476.000	360.000	3.191	22.136	152348.544	146948.039	-3.675