

Continuing Calibration Report for SequenceID : FF080524AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF014394.D

Aliphatic C9-C12	7884586.000	60.000	3.116	6.730	131409.767	130838.329	-0.437
Aliphatic C12-C16	5220008.000	40.000	6.731	10.162	130500.200	129102.248	-1.083
Aliphatic C16-C21	7689538.000	60.000	10.163	13.525	128158.967	126791.817	-1.078
Aliphatic C21-C28	9180483.000	80.000	13.526	17.187	114756.038	114990.989	0.204
Aliphatic C28-C40	9467450.000	120.000	17.188	22.029	78895.417	87817.128	10.159
Aliphatic EPH	39442065.000	360.000	3.116	22.029	109561.292	112108.981	2.273