

Continuing Calibration Report for SequenceID : FG080224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014030.D

Aliphatic C9-C12	7857693.000	60.000	3.116	6.739	130961.550	131889.066	0.703
Aliphatic C12-C16	5199412.000	40.000	6.740	10.183	129985.300	130673.936	0.527
Aliphatic C16-C21	7506750.000	60.000	10.184	13.556	125112.500	125509.453	0.316
Aliphatic C21-C28	9326537.000	80.000	13.557	17.226	116581.713	117202.040	0.529
Aliphatic C28-C40	12777030.000	120.000	17.227	22.108	106475.250	114652.091	7.132
Aliphatic EPH	42667422.000	360.000	3.116	22.108	118520.617	121681.341	2.598