

Continuing Calibration Report for SequenceID : FG080224AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FG014040.D

Aliphatic C9-C12	7784367.000	60.000	3.116	6.739	129739.450	131889.066	1.630
Aliphatic C12-C16	5095425.000	40.000	6.740	10.183	127385.625	130673.936	2.516
Aliphatic C16-C21	7339838.000	60.000	10.184	13.556	122330.633	125509.453	2.533
Aliphatic C21-C28	9203561.000	80.000	13.557	17.226	115044.513	117202.040	1.841
Aliphatic C28-C40	12181496.000	120.000	17.227	22.108	101512.467	114652.091	11.460
Aliphatic EPH	41604687.000	360.000	3.116	22.108	115568.575	121681.341	5.024