

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB073024

Review By	yogesh	Review On	7/30/2024 1:35:29 PM
Supervise By	Ankita	Supervise On	7/31/2024 9:03:24 AM
SubDirectory	FB073024	HP Acquire Method	HP Processing Method FB073024

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP23534,PP23540,PP23541,PP23542,PP23543,PP23544
CCC Internal Standard/PEM	PP23545,PP23546,PP23547
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23535,PP23548

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB030831.D	30 Jul 2024 8:39	Spiking error	YP/AJ	Not Ok
2	10 GRO STD		FB030832.D	30 Jul 2024 9:17		YP/AJ	Ok
3	20 GRO STD		FB030833.D	30 Jul 2024 9:56		YP/AJ	Ok
4	50 GRO STD		FB030834.D	30 Jul 2024 10:35		YP/AJ	Ok
5	100 GRO STD		FB030835.D	30 Jul 2024 11:14		YP/AJ	Ok
6	5 GRO STD		FB030836.D	30 Jul 2024 12:35		YP/AJ	Ok
7	FB073024GROICV		FB030837.D	30 Jul 2024 13:14		YP/AJ	Ok
8	20 PPB GRO STD		FB030838.D	30 Jul 2024 13:53		YP/AJ	Ok
9	VBF0730S1		FB030839.D	30 Jul 2024 14:33		YP/AJ	Ok
10	VBF0730S2		FB030840.D	30 Jul 2024 15:12		YP/AJ	Ok
11	BSF0730S1		FB030841.D	30 Jul 2024 15:51		YP/AJ	Ok
12	P3390-01		FB030842.D	30 Jul 2024 17:10	1 PPB SOIL LOD	YP/AJ	Ok
13	P3390-01		FB030843.D	30 Jul 2024 17:50	4 PPB SOIL LOD	YP/AJ	Ok
14	P3390-02		FB030844.D	30 Jul 2024 18:29	5 PPB SOIL LOQ	YP/AJ	Ok
15	BSF0730S2		FB030845.D	30 Jul 2024 19:48		YP/AJ	Ok
16	20 PPB GRO STD		FB030846.D	30 Jul 2024 20:27		YP/AJ	Ok
17	VBF0730W1		FB030847.D	30 Jul 2024 21:45		YP/AJ	Ok
18	BSF0730W1		FB030848.D	30 Jul 2024 22:24		YP/AJ	Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB073024

Review By	yogesh	Review On	7/30/2024 1:35:29 PM
Supervise By	Ankita	Supervise On	7/31/2024 9:03:24 AM
SubDirectory	FB073024	HP Acquire Method	HP Processing Method
			FB073024
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP23534,PP23540,PP23541,PP23542,PP23543,PP23544		
CCC Internal Standard/PEM	PP23545,PP23546,PP23547		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP23535,PP23548		

19	BSF0730W2		FB030849.D	30 Jul 2024 23:03		YP/AJ	Ok
20	P3390-07		FB030850.D	31 Jul 2024 00:20	1 PPB WATER LOD	YP/AJ	Ok
21	P3390-07		FB030851.D	31 Jul 2024 00:59	4 PPB WATER LOD	YP/AJ	Ok
22	P3390-08		FB030852.D	31 Jul 2024 1:38	5 PPB WATER LOQ	YP/AJ	Ok
23	BSF0730W3		FB030853.D	31 Jul 2024 2:16		YP/AJ	Ok
24	20 PPB GRO STD		FB030854.D	31 Jul 2024 2:55		YP/AJ	Ok

M : Manual Integration