

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD_S Calibration Date(s): 07/12/2024 07/12/2024

Calibration Times: 10:31 12:08

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS027149.D</u>	RT 500 = <u>PS027150.D</u>
	RT 750 = <u>PS027151.D</u>	RT 1000 = <u>PS027152.D</u>
		RT 1500 = <u>PS027153.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-T	9.47	9.47	9.47	9.47	9.47	9.47	9.37	9.57
2,4,5-TP(Silvex)	9.18	9.18	9.18	9.18	9.18	9.18	9.08	9.28
2,4-D	8.31	8.31	8.31	8.31	8.31	8.31	8.21	8.41
2,4-DB	10.04	10.04	10.04	10.04	10.04	10.04	9.94	10.14
2,4-DCAA	7.21	7.20	7.21	7.20	7.21	7.20	7.10	7.30
DICAMBA	7.39	7.39	7.39	7.39	7.39	7.39	7.29	7.49
DICHLORPROP	8.09	8.09	8.09	8.09	8.09	8.09	7.99	8.19
Dinoseb	11.24	11.24	11.24	11.24	11.23	11.23	11.13	11.33



Contract:	<u>CHEM02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>P3390</u>	SAS No.:	<u>P3390</u>	SDG NO.:	<u>P3390</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>07/12/2024</u>	<u>07/12/2024</u>		
		Calibration Times:		<u>10:31</u>	<u>12:08</u>		

LAB FILE ID:	RT 200 =	<u>PS027149.D</u>	RT 500 =	<u>PS027150.D</u>
RT 750 =	<u>PS027151.D</u>	RT 1000 =	<u>PS027152.D</u>	RT 1500 =
				<u>PS027153.D</u>

[illegible]