

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/01/2024 Initial Calibration Date(s): 07/12/2024 07/12/2024

Continuing Calib Time: 19:46 Initial Calibration Time(s): 10:31 12:08

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.38	7.39	7.29	7.49	0.01
2,4-DCAA	7.19	7.21	7.11	7.31	0.02
DICHLORPROP	8.07	8.09	7.99	8.19	0.02
2,4-D	8.30	8.31	8.21	8.41	0.01
2,4,5-TP(Silvex)	9.16	9.18	9.08	9.28	0.02
2,4,5-T	9.45	9.47	9.37	9.57	0.02
2,4-DB	10.02	10.04	9.94	10.14	0.02
Dinoseb	11.21	11.24	11.14	11.34	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/01/2024 Initial Calibration Date(s): 07/12/2024 07/12/2024

Continuing Calib Time: 19:46 Initial Calibration Time(s): 10:31 12:08

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.95	7.95	7.85	8.05	0.00
2,4-DCAA	7.75	7.75	7.65	7.85	0.00
DICHLORPROP	8.66	8.67	8.57	8.77	0.01
2,4-D	8.99	8.99	8.89	9.09	0.00
2,4,5-TP(Silvex)	9.89	9.90	9.80	10.00	0.01
2,4,5-T	10.31	10.32	10.22	10.42	0.01
2,4-DB	10.88	10.88	10.78	10.98	0.00
Dinoseb	11.26	11.26	11.16	11.36	0.00