

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 07/12/2024 07/12/2024

Client Sample No.: CCAL01 Date Analyzed: 08/01/2024

Lab Sample No.: HSTDCCC750 Data File : PS027260.D Time Analyzed: 09:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.452	9.369	9.569	730.460	712.500	2.5
2,4,5-TP(Silvex)	9.164	9.079	9.279	728.540	712.500	2.3
2,4-D	8.298	8.213	8.413	711.990	705.000	1.0
2,4-DB	10.020	9.938	10.138	726.430	712.500	2.0
2,4-DCAA	7.193	7.105	7.305	745.670	750.000	-0.6
DICAMBA	7.376	7.289	7.489	700.120	705.000	-0.7
DICHLORPROP	8.072	7.985	8.185	686.050	705.000	-2.7
Dinoseb	11.214	11.135	11.335	721.030	705.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/12/2024 07/12/2024

Client Sample No.: CCAL01 Date Analyzed: 08/01/2024

Lab Sample No.: HSTDCCC750 Data File : PS027260.D Time Analyzed: 09:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.313	10.216	10.416	695.250	712.500	-2.4
2,4,5-TP(Silvex)	9.894	9.796	9.996	699.500	712.500	-1.8
2,4-D	8.991	8.894	9.094	676.290	705.000	-4.1
2,4-DB	10.877	10.780	10.980	663.910	712.500	-6.8
2,4-DCAA	7.750	7.652	7.852	701.540	750.000	-6.5
DICAMBA	7.950	7.852	8.052	685.920	705.000	-2.7
DICHLORPROP	8.662	8.565	8.765	674.450	705.000	-4.3
Dinoseb	11.255	11.159	11.359	710.860	705.000	0.8