

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 07/12/2024 07/12/2024

Client Sample No.: CCAL02 Date Analyzed: 08/01/2024

Lab Sample No.: HSTDCCC750 Data File : PS027270.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.451	9.369	9.569	746.460	712.500	4.8
2,4,5-TP(Silvex)	9.163	9.079	9.279	740.300	712.500	3.9
2,4-D	8.298	8.213	8.413	731.480	705.000	3.8
2,4-DB	10.020	9.938	10.138	749.810	712.500	5.2
2,4-DCAA	7.192	7.105	7.305	759.430	750.000	1.3
DICAMBA	7.375	7.289	7.489	713.340	705.000	1.2
DICHLORPROP	8.071	7.985	8.185	699.670	705.000	-0.8
Dinoseb	11.213	11.135	11.335	734.550	705.000	4.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/12/2024 07/12/2024

Client Sample No.: CCAL02 Date Analyzed: 08/01/2024

Lab Sample No.: HSTDCCC750 Data File : PS027270.D Time Analyzed: 13:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.312	10.216	10.416	714.190	712.500	0.2
2,4,5-TP(Silvex)	9.894	9.796	9.996	715.390	712.500	0.4
2,4-D	8.992	8.894	9.094	701.010	705.000	-0.6
2,4-DB	10.878	10.780	10.980	695.030	712.500	-2.5
2,4-DCAA	7.751	7.652	7.852	729.250	750.000	-2.8
DICAMBA	7.950	7.852	8.052	712.820	705.000	1.1
DICHLORPROP	8.663	8.565	8.765	705.780	705.000	0.1
Dinoseb	11.256	11.159	11.359	735.030	705.000	4.3