



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

LOD-MDL-WATER-01-Q

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Lab Sample ID: P3390-07 Date(s) Analyzed: 08/06/2024 08/06/2024

Instrument ID (1): ECD_R Instrument ID (2): ECD_R

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PR068003.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.902	4.852	4.952	0.23	0.21	11.06
COLUMN 1	2	4.925	4.875	4.975	0.19		
	3	4.99	4.94	5.04	0.20		
	4	5.091	5.041	5.141	0.19		
	5	5.404	5.354	5.454	0.23		
COLUMN 2	1	4.144	4.094	4.194	0.23	0.23	
	2	4.163	4.113	4.213	0.23		
	3	4.344	4.294	4.394	0.23		
	4	4.396	4.346	4.446	0.26		
	5	4.615	4.565	4.665	0.22		
Aroclor-1260	1	6.616	6.566	6.666	0.22	0.25	5.61
COLUMN 1	2	6.899	6.849	6.949	0.24		
	3	7.289	7.239	7.339	0.26		
	4	7.531	7.481	7.581	0.25		
	5	7.867	7.817	7.917	0.25		
COLUMN 2	1	5.72	5.67	5.77	0.25	0.26	
	2	5.926	5.876	5.976	0.26		
	3	6.087	6.037	6.137	0.29		
	4	6.597	6.547	6.647	0.26		
	5	6.863	6.813	6.913	0.25		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

LOD-MDL-WATER-01-Q

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Lab Sample ID: P3390-07 Date(s) Analyzed: 08/09/2024 08/09/2024

Instrument ID (1): ECD_R Instrument ID (2): ECD_R

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PR068110.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1221	1	3.891	3.841	3.941	0.28		
	2	3.988	3.938	4.038	0.46		
	3	4.066	4.016	4.116	0.22		
	4	0	0	0	0		
	5	0	0	0	0		
COLUMN 1	1					0.32	
	2						
	3						
	4						
	5						
COLUMN 2	1	3.235	3.185	3.285	0.36		
	2	3.319	3.269	3.369	0.50		
	3	3.4	3.35	3.45	0.22		
	4	0	0	0	0		
	5	0	0	0	0		
						0.36	12.52