

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/12/2024 Initial Calibration Date(s): 08/09/2024 08/10/2024

Continuing Calib Time: 09:46 Initial Calibration Time(s): 18:05 01:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1016-2	(2)	5.62	5.62	5.52	5.72	0.00
Aroclor-1016-3	(3)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-4	(4)	5.79	5.79	5.69	5.89	0.00
Aroclor-1016-5	(5)	6.08	6.08	5.98	6.18	0.00
Aroclor-1260-1	(1)	7.22	7.22	7.12	7.32	0.00
Aroclor-1260-2	(2)	7.48	7.48	7.38	7.58	0.00
Aroclor-1260-3	(3)	7.84	7.84	7.74	7.94	0.00
Aroclor-1260-4	(4)	8.07	8.07	7.97	8.17	0.01
Aroclor-1260-5	(5)	8.39	8.39	8.29	8.49	0.00
Tetrachloro-m-xylene		4.43	4.43	4.33	4.53	0.00
Decachlorobiphenyl		10.22	10.22	10.12	10.32	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/12/2024 Initial Calibration Date(s): 08/09/2024 08/10/2024

Continuing Calib Time: 09:46 Initial Calibration Time(s): 18:05 01:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.59	4.59	4.49	4.69	0.00
Aroclor-1016-2	(2)	4.61	4.61	4.51	4.71	0.00
Aroclor-1016-3	(3)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-4	(4)	4.82	4.82	4.72	4.92	0.00
Aroclor-1016-5	(5)	5.03	5.03	4.93	5.13	0.00
Aroclor-1260-1	(1)	6.05	6.05	5.95	6.15	0.00
Aroclor-1260-2	(2)	6.23	6.23	6.13	6.33	0.00
Aroclor-1260-3	(3)	6.38	6.38	6.28	6.48	0.00
Aroclor-1260-4	(4)	6.85	6.85	6.75	6.95	0.00
Aroclor-1260-5	(5)	7.09	7.09	6.99	7.19	0.00
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00
Decachlorobiphenyl		8.43	8.43	8.33	8.53	0.00