

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/12/2024 Initial Calibration Date(s): 08/09/2024 08/10/2024

Continuing Calib Time: 13:48 Initial Calibration Time(s): 18:05 01:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.46	6.46	6.36	6.56	0.00
Aroclor-1254-2	(2)	6.68	6.68	6.58	6.78	0.00
Aroclor-1254-3	(3)	7.05	7.05	6.95	7.15	0.00
Aroclor-1254-4	(4)	7.34	7.34	7.24	7.44	0.00
Aroclor-1254-5	(5)	7.76	7.76	7.66	7.86	0.00
Tetrachloro-m-xylene		4.43	4.43	4.33	4.53	0.00
Decachlorobiphenyl		10.22	10.22	10.12	10.32	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/12/2024 Initial Calibration Date(s): 08/09/2024 08/10/2024

Continuing Calib Time: 13:48 Initial Calibration Time(s): 18:05 01:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.38	5.38	5.28	5.48	0.00
Aroclor-1254-2 (2)	5.52	5.52	5.42	5.62	0.00
Aroclor-1254-3 (3)	5.92	5.92	5.82	6.02	0.00
Aroclor-1254-4 (4)	6.14	6.14	6.04	6.24	0.00
Aroclor-1254-5 (5)	6.55	6.55	6.45	6.65	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
Decachlorobiphenyl	8.43	8.43	8.33	8.53	0.00