

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/21/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 00:00 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.88	5.88	5.78	5.98	0.00
Aroclor-1016-2	(2)	5.90	5.90	5.80	6.00	0.00
Aroclor-1016-3	(3)	5.96	5.97	5.87	6.07	0.01
Aroclor-1016-4	(4)	6.06	6.06	5.96	6.16	0.00
Aroclor-1016-5	(5)	6.36	6.36	6.26	6.46	0.00
Aroclor-1260-1	(1)	7.48	7.49	7.39	7.59	0.01
Aroclor-1260-2	(2)	7.74	7.74	7.64	7.84	0.00
Aroclor-1260-3	(3)	8.10	8.10	8.00	8.20	0.00
Aroclor-1260-4	(4)	8.34	8.34	8.24	8.44	0.00
Aroclor-1260-5	(5)	8.68	8.68	8.58	8.78	0.00
Tetrachloro-m-xylene		4.70	4.71	4.61	4.81	0.01
Decachlorobiphenyl		10.62	10.63	10.53	10.73	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/21/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 00:00 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-2	(2)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-3	(3)	5.29	5.29	5.19	5.39	0.00
Aroclor-1016-4	(4)	5.33	5.33	5.23	5.43	0.00
Aroclor-1016-5	(5)	5.55	5.55	5.45	5.65	0.00
Aroclor-1260-1	(1)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-2	(2)	6.79	6.79	6.69	6.89	0.00
Aroclor-1260-3	(3)	6.94	6.95	6.85	7.05	0.01
Aroclor-1260-4	(4)	7.42	7.42	7.32	7.52	0.00
Aroclor-1260-5	(5)	7.66	7.66	7.56	7.76	0.00
Tetrachloro-m-xylene		3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl		9.14	9.15	9.05	9.25	0.01