

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/22/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 09:09 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.88	5.88	5.78	5.98	0.00
Aroclor-1242-2 (2)	5.91	5.90	5.80	6.00	-0.01
Aroclor-1242-3 (3)	5.97	5.96	5.86	6.06	-0.01
Aroclor-1242-4 (4)	6.07	6.06	5.96	6.16	-0.01
Aroclor-1242-5 (5)	6.81	6.80	6.70	6.90	-0.01
Tetrachloro-m-xylene	4.71	4.70	4.60	4.80	-0.01
Decachlorobiphenyl	10.63	10.63	10.53	10.73	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/22/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 09:09 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.09	5.09	4.99	5.19	0.00
Aroclor-1242-2 (2)	5.11	5.11	5.01	5.21	0.00
Aroclor-1242-3 (3)	5.30	5.29	5.19	5.39	-0.01
Aroclor-1242-4 (4)	5.38	5.37	5.27	5.47	-0.01
Aroclor-1242-5 (5)	5.91	5.91	5.81	6.01	0.00
Tetrachloro-m-xylene	3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl	9.15	9.15	9.05	9.25	0.00