

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/22/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 09:26 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.88	5.88	5.78	5.98	0.00
Aroclor-1248-2	(2)	6.15	6.15	6.05	6.25	0.00
Aroclor-1248-3	(3)	6.36	6.36	6.26	6.46	0.00
Aroclor-1248-4	(4)	6.76	6.76	6.66	6.86	0.00
Aroclor-1248-5	(5)	6.80	6.80	6.70	6.90	0.00
Tetrachloro-m-xylene		4.70	4.70	4.60	4.80	0.00
Decachlorobiphenyl		10.62	10.63	10.53	10.73	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/22/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 09:26 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	5.09	5.09	4.99	5.19	0.00
Aroclor-1248-2	(2)	5.33	5.33	5.23	5.43	0.00
Aroclor-1248-3	(3)	5.37	5.37	5.27	5.47	0.00
Aroclor-1248-4	(4)	5.55	5.55	5.45	5.65	0.00
Aroclor-1248-5	(5)	5.95	5.95	5.85	6.05	0.00
Tetrachloro-m-xylene		3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl		9.14	9.15	9.05	9.25	0.01