

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/22/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 10:38 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1254-1	(1)	6.74	6.74	6.64	6.84	0.00
Aroclor-1254-2	(2)	6.95	6.96	6.86	7.06	0.01
Aroclor-1254-3	(3)	7.32	7.32	7.22	7.42	0.00
Aroclor-1254-4	(4)	7.61	7.61	7.51	7.71	0.00
Aroclor-1254-5	(5)	8.03	8.03	7.93	8.13	0.01
Tetrachloro-m-xylene		4.70	4.71	4.61	4.81	0.01
Decachlorobiphenyl		10.63	10.63	10.53	10.73	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3390 **SAS No.:** P3390 **SDG NO.:** P3390

Continuing Calib Date: 08/22/2024 **Initial Calibration Date(s):** 08/20/2024 08/20/2024

Continuing Calib Time: 10:38 **Initial Calibration Time(s):** 11:41 19:15

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.91	5.91	5.81	6.01	0.00
Aroclor-1254-2 (2)	6.06	6.06	5.96	6.16	0.00
Aroclor-1254-3 (3)	6.47	6.47	6.37	6.57	0.00
Aroclor-1254-4 (4)	6.70	6.70	6.60	6.80	0.00
Aroclor-1254-5 (5)	7.12	7.12	7.02	7.22	0.00
Tetrachloro-m-xylene	3.98	3.98	3.88	4.08	0.00
Decachlorobiphenyl	9.15	9.15	9.05	9.25	0.00