

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/23/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 02:29 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.88       | 5.88      | 5.78                   | 5.98  | 0.00       |
| Aroclor-1016-2       | (2) | 5.90       | 5.90      | 5.80                   | 6.00  | 0.00       |
| Aroclor-1016-3       | (3) | 5.96       | 5.97      | 5.87                   | 6.07  | 0.01       |
| Aroclor-1016-4       | (4) | 6.06       | 6.06      | 5.96                   | 6.16  | 0.00       |
| Aroclor-1016-5       | (5) | 6.36       | 6.36      | 6.26                   | 6.46  | 0.00       |
| Aroclor-1260-1       | (1) | 7.48       | 7.49      | 7.39                   | 7.59  | 0.01       |
| Aroclor-1260-2       | (2) | 7.74       | 7.74      | 7.64                   | 7.84  | 0.00       |
| Aroclor-1260-3       | (3) | 8.10       | 8.10      | 8.00                   | 8.20  | 0.00       |
| Aroclor-1260-4       | (4) | 8.34       | 8.34      | 8.24                   | 8.44  | 0.00       |
| Aroclor-1260-5       | (5) | 8.68       | 8.68      | 8.58                   | 8.78  | 0.00       |
| Tetrachloro-m-xylene |     | 4.70       | 4.71      | 4.61                   | 4.81  | 0.01       |
| Decachlorobiphenyl   |     | 10.62      | 10.63     | 10.53                  | 10.73 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/23/2024 Initial Calibration Date(s): 08/20/2024 08/20/2024

Continuing Calib Time: 02:29 Initial Calibration Time(s): 11:41 19:15

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 5.09       | 5.09      | 4.99      | 5.19 | 0.00       |
| Aroclor-1016-2       | (2) | 5.11       | 5.11      | 5.01      | 5.21 | 0.00       |
| Aroclor-1016-3       | (3) | 5.29       | 5.29      | 5.19      | 5.39 | 0.00       |
| Aroclor-1016-4       | (4) | 5.33       | 5.33      | 5.23      | 5.43 | 0.00       |
| Aroclor-1016-5       | (5) | 5.55       | 5.55      | 5.45      | 5.65 | 0.00       |
| Aroclor-1260-1       | (1) | 6.60       | 6.60      | 6.50      | 6.70 | 0.00       |
| Aroclor-1260-2       | (2) | 6.79       | 6.79      | 6.69      | 6.89 | 0.00       |
| Aroclor-1260-3       | (3) | 6.95       | 6.95      | 6.85      | 7.05 | 0.00       |
| Aroclor-1260-4       | (4) | 7.42       | 7.42      | 7.32      | 7.52 | 0.00       |
| Aroclor-1260-5       | (5) | 7.66       | 7.66      | 7.56      | 7.76 | 0.00       |
| Tetrachloro-m-xylene |     | 3.99       | 3.98      | 3.88      | 4.08 | 0.00       |
| Decachlorobiphenyl   |     | 9.14       | 9.15      | 9.05      | 9.25 | 0.01       |