

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 08/05/2024 08/05/2024

Continuing Calib Time: 22:23 Initial Calibration Time(s): 13:01 19:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.57	4.57	4.47	4.67	0.00
Aroclor-1016-2	(2)	4.59	4.59	4.49	4.69	0.00
Aroclor-1016-3	(3)	4.65	4.65	4.55	4.75	0.00
Aroclor-1016-4	(4)	4.74	4.75	4.65	4.85	0.01
Aroclor-1016-5	(5)	5.03	5.04	4.94	5.14	0.01
Aroclor-1260-1	(1)	6.15	6.16	6.06	6.26	0.01
Aroclor-1260-2	(2)	6.42	6.42	6.32	6.52	0.00
Aroclor-1260-3	(3)	6.78	6.78	6.68	6.88	0.00
Aroclor-1260-4	(4)	7.00	7.00	6.90	7.10	0.00
Aroclor-1260-5	(5)	7.32	7.32	7.22	7.42	0.00
Tetrachloro-m-xylene		3.46	3.46	3.36	3.56	0.00
Decachlorobiphenyl		8.73	8.74	8.64	8.84	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 08/05/2024 08/05/2024

Continuing Calib Time: 22:23 Initial Calibration Time(s): 13:01 19:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	3.84	3.84	3.74	3.94	0.00
Aroclor-1016-2	(2)	3.85	3.85	3.75	3.95	0.00
Aroclor-1016-3	(3)	4.01	4.01	3.91	4.11	0.00
Aroclor-1016-4	(4)	4.06	4.06	3.96	4.16	0.00
Aroclor-1016-5	(5)	4.26	4.26	4.16	4.36	0.00
Aroclor-1260-1	(1)	5.26	5.26	5.16	5.36	0.00
Aroclor-1260-2	(2)	5.45	5.45	5.35	5.55	0.00
Aroclor-1260-3	(3)	5.59	5.59	5.49	5.69	0.00
Aroclor-1260-4	(4)	6.05	6.05	5.95	6.15	0.00
Aroclor-1260-5	(5)	6.29	6.30	6.20	6.40	0.01
Tetrachloro-m-xylene		2.83	2.83	2.73	2.93	0.00
Decachlorobiphenyl		7.61	7.61	7.51	7.71	0.00