

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/07/2024 Initial Calibration Date(s): 08/05/2024 08/05/2024

Continuing Calib Time: 02:46 Initial Calibration Time(s): 19:46 22:27

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.92	4.90	4.80	5.00	-0.02
Aroclor-1016-2	(2)	4.94	4.92	4.82	5.02	-0.02
Aroclor-1016-3	(3)	5.00	4.99	4.89	5.09	-0.01
Aroclor-1016-4	(4)	5.11	5.09	4.99	5.19	-0.02
Aroclor-1016-5	(5)	5.42	5.41	5.31	5.51	-0.01
Aroclor-1260-1	(1)	6.63	6.62	6.52	6.72	-0.01
Aroclor-1260-2	(2)	6.92	6.90	6.80	7.00	-0.02
Aroclor-1260-3	(3)	7.31	7.29	7.19	7.39	-0.02
Aroclor-1260-4	(4)	7.55	7.53	7.43	7.63	-0.02
Aroclor-1260-5	(5)	7.89	7.87	7.77	7.97	-0.02
Tetrachloro-m-xylene		3.69	3.68	3.58	3.78	-0.01
Decachlorobiphenyl		9.61	9.58	9.48	9.68	-0.03

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/07/2024 Initial Calibration Date(s): 08/05/2024 08/05/2024

Continuing Calib Time: 02:46 Initial Calibration Time(s): 19:46 22:27

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.14	4.14	4.04	4.24	0.00
Aroclor-1016-2	(2)	4.16	4.16	4.06	4.26	0.00
Aroclor-1016-3	(3)	4.34	4.34	4.24	4.44	0.00
Aroclor-1016-4	(4)	4.39	4.40	4.30	4.50	0.01
Aroclor-1016-5	(5)	4.61	4.62	4.52	4.72	0.01
Aroclor-1260-1	(1)	5.72	5.72	5.62	5.82	0.00
Aroclor-1260-2	(2)	5.93	5.93	5.83	6.03	0.00
Aroclor-1260-3	(3)	6.09	6.09	5.99	6.19	0.00
Aroclor-1260-4	(4)	6.60	6.60	6.50	6.70	0.00
Aroclor-1260-5	(5)	6.87	6.86	6.76	6.96	-0.01
Tetrachloro-m-xylene		3.01	3.01	2.91	3.11	0.00
Decachlorobiphenyl		8.34	8.34	8.24	8.44	0.00