

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/09/2024 08/09/2024

Client Sample No.: CCAL01 Date Analyzed: 08/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO105433.D Time Analyzed: 19:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.597	5.501	5.701	486.750	500.000	-2.7
Aroclor-1016-2	5.621	5.523	5.723	477.750	500.000	-4.5
Aroclor-1016-3	5.683	5.585	5.785	484.670	500.000	-3.1
Aroclor-1016-4	5.783	5.686	5.886	475.400	500.000	-4.9
Aroclor-1016-5	6.079	5.982	6.182	468.710	500.000	-6.3
Aroclor-1260-1	7.215	7.117	7.317	466.940	500.000	-6.6
Aroclor-1260-2	7.473	7.375	7.575	486.430	500.000	-2.7
Aroclor-1260-3	7.836	7.738	7.938	532.040	500.000	6.4
Aroclor-1260-4	8.063	7.966	8.166	513.380	500.000	2.7
Aroclor-1260-5	8.386	8.289	8.489	503.180	500.000	0.6
Decachlorobiphenyl	10.217	10.123	10.323	51.690	50.000	3.4
Tetrachloro-m-xylene	4.427	4.329	4.529	48.030	50.000	-3.9

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/09/2024 08/09/2024

Client Sample No.: CCAL01 Date Analyzed: 08/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO105433.D Time Analyzed: 19:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.590	4.492	4.692	478.970	500.000	-4.2
Aroclor-1016-2	4.607	4.510	4.710	482.000	500.000	-3.6
Aroclor-1016-3	4.780	4.681	4.881	486.160	500.000	-2.8
Aroclor-1016-4	4.821	4.723	4.923	480.770	500.000	-3.8
Aroclor-1016-5	5.029	4.932	5.132	487.050	500.000	-2.6
Aroclor-1260-1	6.044	5.946	6.146	478.370	500.000	-4.3
Aroclor-1260-2	6.230	6.132	6.332	469.550	500.000	-6.1
Aroclor-1260-3	6.380	6.282	6.482	454.620	500.000	-9.1
Aroclor-1260-4	6.846	6.748	6.948	483.360	500.000	-3.3
Aroclor-1260-5	7.086	6.988	7.188	478.280	500.000	-4.3
Decachlorobiphenyl	8.427	8.328	8.528	48.380	50.000	-3.2
Tetrachloro-m-xylene	3.536	3.437	3.637	48.480	50.000	-3.0