

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/09/2024 08/09/2024

Client Sample No.: CCAL02 Date Analyzed: 08/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO105444.D Time Analyzed: 22:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.598	5.501	5.701	471.690	500.000	-5.7
Aroclor-1016-2	5.620	5.523	5.723	472.870	500.000	-5.4
Aroclor-1016-3	5.683	5.585	5.785	474.660	500.000	-5.1
Aroclor-1016-4	5.783	5.686	5.886	467.320	500.000	-6.5
Aroclor-1016-5	6.079	5.982	6.182	460.760	500.000	-7.8
Aroclor-1260-1	7.215	7.117	7.317	473.820	500.000	-5.2
Aroclor-1260-2	7.473	7.375	7.575	483.500	500.000	-3.3
Aroclor-1260-3	7.835	7.738	7.938	518.300	500.000	3.7
Aroclor-1260-4	8.063	7.966	8.166	503.130	500.000	0.6
Aroclor-1260-5	8.385	8.289	8.489	503.410	500.000	0.7
Decachlorobiphenyl	10.216	10.123	10.323	50.930	50.000	1.9
Tetrachloro-m-xylene	4.427	4.329	4.529	47.130	50.000	-5.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/09/2024 08/09/2024

Client Sample No.: CCAL02 Date Analyzed: 08/12/2024

Lab Sample No.: AR1660CCC500 Data File : PO105444.D Time Analyzed: 22:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.590	4.492	4.692	485.670	500.000	-2.9
Aroclor-1016-2	4.608	4.510	4.710	487.230	500.000	-2.6
Aroclor-1016-3	4.780	4.681	4.881	491.430	500.000	-1.7
Aroclor-1016-4	4.821	4.723	4.923	479.830	500.000	-4.0
Aroclor-1016-5	5.029	4.932	5.132	494.660	500.000	-1.1
Aroclor-1260-1	6.044	5.946	6.146	486.700	500.000	-2.7
Aroclor-1260-2	6.230	6.132	6.332	477.660	500.000	-4.5
Aroclor-1260-3	6.381	6.282	6.482	462.930	500.000	-7.4
Aroclor-1260-4	6.847	6.748	6.948	490.960	500.000	-1.8
Aroclor-1260-5	7.087	6.988	7.188	487.870	500.000	-2.4
Decachlorobiphenyl	8.427	8.328	8.528	49.690	50.000	-0.6
Tetrachloro-m-xylene	3.535	3.437	3.637	49.160	50.000	-1.7