

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/20/2024 08/20/2024

Client Sample No.: CCAL01 Date Analyzed: 08/20/2024

Lab Sample No.: AR1660CCC500 Data File : PP066696.D Time Analyzed: 20:55

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.876	5.779	5.979	503.100	500.000	0.6
Aroclor-1016-2	5.899	5.802	6.002	518.110	500.000	3.6
Aroclor-1016-3	5.962	5.865	6.065	520.930	500.000	4.2
Aroclor-1016-4	6.061	5.963	6.163	527.290	500.000	5.5
Aroclor-1016-5	6.356	6.259	6.459	523.350	500.000	4.7
Aroclor-1260-1	7.484	7.387	7.587	522.290	500.000	4.5
Aroclor-1260-2	7.740	7.642	7.842	502.080	500.000	0.4
Aroclor-1260-3	8.100	8.002	8.202	533.500	500.000	6.7
Aroclor-1260-4	8.337	8.241	8.441	515.990	500.000	3.2
Aroclor-1260-5	8.675	8.580	8.780	503.390	500.000	0.7
Decachlorobiphenyl	10.621	10.526	10.726	51.370	50.000	2.7
Tetrachloro-m-xylene	4.702	4.605	4.805	52.520	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/20/2024 08/20/2024

Client Sample No.: CCAL01 Date Analyzed: 08/20/2024

Lab Sample No.: AR1660CCC500 Data File : PP066696.D Time Analyzed: 20:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.085	4.987	5.187	489.010	500.000	-2.2
Aroclor-1016-2	5.105	5.006	5.206	494.980	500.000	-1.0
Aroclor-1016-3	5.287	5.189	5.389	503.320	500.000	0.7
Aroclor-1016-4	5.326	5.227	5.427	495.950	500.000	-0.8
Aroclor-1016-5	5.546	5.448	5.648	497.770	500.000	-0.4
Aroclor-1260-1	6.597	6.499	6.699	489.180	500.000	-2.2
Aroclor-1260-2	6.785	6.687	6.887	487.370	500.000	-2.5
Aroclor-1260-3	6.944	6.846	7.046	477.760	500.000	-4.4
Aroclor-1260-4	7.421	7.323	7.523	485.430	500.000	-2.9
Aroclor-1260-5	7.662	7.564	7.764	487.710	500.000	-2.5
Decachlorobiphenyl	9.144	9.046	9.246	49.590	50.000	-0.8
Tetrachloro-m-xylene	3.978	3.880	4.080	52.230	50.000	4.5