

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/20/2024 08/20/2024

Client Sample No.: CCAL02 Date Analyzed: 08/21/2024

Lab Sample No.: AR1660CCC500 Data File : PP066718.D Time Analyzed: 03:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.877	5.779	5.979	526.300	500.000	5.3
Aroclor-1016-2	5.899	5.802	6.002	526.620	500.000	5.3
Aroclor-1016-3	5.963	5.865	6.065	530.330	500.000	6.1
Aroclor-1016-4	6.062	5.963	6.163	539.990	500.000	8.0
Aroclor-1016-5	6.357	6.259	6.459	546.940	500.000	9.4
Aroclor-1260-1	7.484	7.387	7.587	547.010	500.000	9.4
Aroclor-1260-2	7.740	7.642	7.842	530.560	500.000	6.1
Aroclor-1260-3	8.100	8.002	8.202	559.760	500.000	12.0
Aroclor-1260-4	8.338	8.241	8.441	549.630	500.000	9.9
Aroclor-1260-5	8.675	8.580	8.780	542.700	500.000	8.5
Decachlorobiphenyl	10.622	10.526	10.726	54.270	50.000	8.5
Tetrachloro-m-xylene	4.703	4.605	4.805	53.350	50.000	6.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/20/2024 08/20/2024

Client Sample No.: CCAL02 Date Analyzed: 08/21/2024

Lab Sample No.: AR1660CCC500 Data File : PP066718.D Time Analyzed: 03:04

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.086	4.987	5.187	466.060	500.000	-6.8
Aroclor-1016-2	5.105	5.006	5.206	472.450	500.000	-5.5
Aroclor-1016-3	5.288	5.189	5.389	479.840	500.000	-4.0
Aroclor-1016-4	5.327	5.227	5.427	482.680	500.000	-3.5
Aroclor-1016-5	5.547	5.448	5.648	477.510	500.000	-4.5
Aroclor-1260-1	6.598	6.499	6.699	467.330	500.000	-6.5
Aroclor-1260-2	6.785	6.687	6.887	465.120	500.000	-7.0
Aroclor-1260-3	6.944	6.846	7.046	458.910	500.000	-8.2
Aroclor-1260-4	7.421	7.323	7.523	468.590	500.000	-6.3
Aroclor-1260-5	7.661	7.564	7.764	478.650	500.000	-4.3
Decachlorobiphenyl	9.142	9.046	9.246	48.930	50.000	-2.1
Tetrachloro-m-xylene	3.979	3.880	4.080	49.960	50.000	-0.1