

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/20/2024 08/20/2024

Client Sample No.: CCAL02 Date Analyzed: 08/21/2024

Lab Sample No.: AR1660CCC500 Data File : PP066762.D Time Analyzed: 15:33

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.876	5.779	5.979	491.090	500.000	-1.8
Aroclor-1016-2	5.899	5.802	6.002	511.180	500.000	2.2
Aroclor-1016-3	5.963	5.865	6.065	509.550	500.000	1.9
Aroclor-1016-4	6.062	5.963	6.163	523.970	500.000	4.8
Aroclor-1016-5	6.357	6.259	6.459	551.040	500.000	10.2
Aroclor-1260-1	7.486	7.387	7.587	533.840	500.000	6.8
Aroclor-1260-2	7.741	7.642	7.842	519.240	500.000	3.8
Aroclor-1260-3	8.101	8.002	8.202	544.650	500.000	8.9
Aroclor-1260-4	8.338	8.241	8.441	534.470	500.000	6.9
Aroclor-1260-5	8.676	8.580	8.780	530.450	500.000	6.1
Decachlorobiphenyl	10.625	10.526	10.726	53.870	50.000	7.7
Tetrachloro-m-xylene	4.703	4.605	4.805	50.000	50.000	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/20/2024 08/20/2024

Client Sample No.: CCAL02 Date Analyzed: 08/21/2024

Lab Sample No.: AR1660CCC500 Data File : PP066762.D Time Analyzed: 15:33

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.088	4.987	5.187	413.440	500.000	-17.3
Aroclor-1016-2	5.108	5.006	5.206	414.100	500.000	-17.2
Aroclor-1016-3	5.291	5.189	5.389	423.650	500.000	-15.3
Aroclor-1016-4	5.329	5.227	5.427	431.820	500.000	-13.6
Aroclor-1016-5	5.550	5.448	5.648	440.840	500.000	-11.8
Aroclor-1260-1	6.600	6.499	6.699	424.420	500.000	-15.1
Aroclor-1260-2	6.788	6.687	6.887	420.890	500.000	-15.8
Aroclor-1260-3	6.947	6.846	7.046	409.680	500.000	-18.1
Aroclor-1260-4	7.423	7.323	7.523	420.250	500.000	-16.0
Aroclor-1260-5	7.664	7.564	7.764	432.320	500.000	-13.5
Decachlorobiphenyl	9.146	9.046	9.246	45.780	50.000	-8.4
Tetrachloro-m-xylene	3.982	3.880	4.080	43.110	50.000	-13.8