

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL01 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PQ067657.D Time Analyzed: 08:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.583	4.474	4.674	517.500	500.000	3.5
Aroclor-1016-2	4.602	4.494	4.694	542.460	500.000	8.5
Aroclor-1016-3	4.660	4.552	4.752	546.610	500.000	9.3
Aroclor-1016-4	4.753	4.645	4.845	546.440	500.000	9.3
Aroclor-1016-5	5.045	4.936	5.136	550.330	500.000	10.1
Aroclor-1260-1	6.169	6.055	6.255	522.630	500.000	4.5
Aroclor-1260-2	6.436	6.321	6.521	511.010	500.000	2.2
Aroclor-1260-3	6.797	6.681	6.881	522.580	500.000	4.5
Aroclor-1260-4	7.018	6.901	7.101	525.480	500.000	5.1
Aroclor-1260-5	7.341	7.221	7.421	534.080	500.000	6.8
Decachlorobiphenyl	8.763	8.636	8.836	54.480	50.000	9.0
Tetrachloro-m-xylene	3.465	3.361	3.561	52.950	50.000	5.9

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL01 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PQ067657.D Time Analyzed: 08:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.836	3.737	3.937	539.150	500.000	7.8
Aroclor-1016-2	3.852	3.753	3.953	531.890	500.000	6.4
Aroclor-1016-3	4.014	3.914	4.114	564.680	500.000	12.9
Aroclor-1016-4	4.061	3.962	4.162	557.310	500.000	11.5
Aroclor-1016-5	4.257	4.158	4.358	543.460	500.000	8.7
Aroclor-1260-1	5.255	5.157	5.357	533.230	500.000	6.6
Aroclor-1260-2	5.445	5.346	5.546	526.080	500.000	5.2
Aroclor-1260-3	5.588	5.488	5.688	534.130	500.000	6.8
Aroclor-1260-4	6.050	5.951	6.151	528.900	500.000	5.8
Aroclor-1260-5	6.295	6.196	6.396	527.210	500.000	5.4
Decachlorobiphenyl	7.612	7.512	7.712	52.500	50.000	5.0
Tetrachloro-m-xylene	2.828	2.729	2.929	56.410	50.000	12.8