

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL02 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PQ067669.D Time Analyzed: 11:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.573	4.474	4.674	512.400	500.000	2.5
Aroclor-1016-2	4.593	4.494	4.694	508.450	500.000	1.7
Aroclor-1016-3	4.651	4.552	4.752	510.270	500.000	2.1
Aroclor-1016-4	4.744	4.645	4.845	499.820	500.000	0.0
Aroclor-1016-5	5.034	4.936	5.136	515.350	500.000	3.1
Aroclor-1260-1	6.154	6.055	6.255	510.830	500.000	2.2
Aroclor-1260-2	6.420	6.321	6.521	497.140	500.000	-0.6
Aroclor-1260-3	6.779	6.681	6.881	498.610	500.000	-0.3
Aroclor-1260-4	7.000	6.901	7.101	495.900	500.000	-0.8
Aroclor-1260-5	7.319	7.221	7.421	507.780	500.000	1.6
Decachlorobiphenyl	8.732	8.636	8.836	53.070	50.000	6.1
Tetrachloro-m-xylene	3.460	3.361	3.561	52.640	50.000	5.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL02 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PQ067669.D Time Analyzed: 11:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	3.836	3.737	3.937	517.150	500.000	3.4
Aroclor-1016-2	3.852	3.753	3.953	533.010	500.000	6.6
Aroclor-1016-3	4.014	3.914	4.114	534.700	500.000	6.9
Aroclor-1016-4	4.062	3.962	4.162	524.080	500.000	4.8
Aroclor-1016-5	4.258	4.158	4.358	520.600	500.000	4.1
Aroclor-1260-1	5.256	5.157	5.357	511.190	500.000	2.2
Aroclor-1260-2	5.446	5.346	5.546	503.310	500.000	0.7
Aroclor-1260-3	5.588	5.488	5.688	515.270	500.000	3.1
Aroclor-1260-4	6.050	5.951	6.151	511.770	500.000	2.4
Aroclor-1260-5	6.296	6.196	6.396	507.260	500.000	1.5
Decachlorobiphenyl	7.610	7.512	7.712	50.660	50.000	1.3
Tetrachloro-m-xylene	2.828	2.729	2.929	52.210	50.000	4.4