

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL01 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PR068000.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.904	4.802	5.002	521.970	500.000	4.4
Aroclor-1016-2	4.926	4.824	5.024	513.010	500.000	2.6
Aroclor-1016-3	4.990	4.888	5.088	514.430	500.000	2.9
Aroclor-1016-4	5.093	4.991	5.191	511.080	500.000	2.2
Aroclor-1016-5	5.407	5.306	5.506	508.140	500.000	1.6
Aroclor-1260-1	6.619	6.518	6.718	497.270	500.000	-0.5
Aroclor-1260-2	6.902	6.800	7.000	491.640	500.000	-1.7
Aroclor-1260-3	7.291	7.190	7.390	495.520	500.000	-0.9
Aroclor-1260-4	7.534	7.432	7.632	500.930	500.000	0.2
Aroclor-1260-5	7.872	7.769	7.969	511.120	500.000	2.2
Decachlorobiphenyl	9.586	9.483	9.683	50.480	50.000	1.0
Tetrachloro-m-xylene	3.680	3.578	3.778	52.050	50.000	4.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL01 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PR068000.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.143	4.044	4.244	533.070	500.000	6.6
Aroclor-1016-2	4.162	4.063	4.263	533.090	500.000	6.6
Aroclor-1016-3	4.343	4.244	4.444	551.750	500.000	10.4
Aroclor-1016-4	4.394	4.296	4.496	546.330	500.000	9.3
Aroclor-1016-5	4.614	4.516	4.716	550.520	500.000	10.1
Aroclor-1260-1	5.719	5.621	5.821	527.540	500.000	5.5
Aroclor-1260-2	5.926	5.827	6.027	529.090	500.000	5.8
Aroclor-1260-3	6.086	5.988	6.188	537.840	500.000	7.6
Aroclor-1260-4	6.597	6.499	6.699	544.800	500.000	9.0
Aroclor-1260-5	6.863	6.764	6.964	532.660	500.000	6.5
Decachlorobiphenyl	8.335	8.235	8.435	50.880	50.000	1.8
Tetrachloro-m-xylene	3.011	2.911	3.111	50.800	50.000	1.6