

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL02 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PR068012.D Time Analyzed: 12:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.902	4.802	5.002	512.140	500.000	2.4
Aroclor-1016-2	4.924	4.824	5.024	514.300	500.000	2.9
Aroclor-1016-3	4.988	4.888	5.088	505.980	500.000	1.2
Aroclor-1016-4	5.091	4.991	5.191	503.110	500.000	0.6
Aroclor-1016-5	5.406	5.306	5.506	490.830	500.000	-1.8
Aroclor-1260-1	6.617	6.518	6.718	514.100	500.000	2.8
Aroclor-1260-2	6.900	6.800	7.000	508.500	500.000	1.7
Aroclor-1260-3	7.289	7.190	7.390	501.920	500.000	0.4
Aroclor-1260-4	7.530	7.432	7.632	510.730	500.000	2.1
Aroclor-1260-5	7.867	7.769	7.969	528.340	500.000	5.7
Decachlorobiphenyl	9.583	9.483	9.683	52.720	50.000	5.4
Tetrachloro-m-xylene	3.678	3.578	3.778	50.870	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/05/2024 08/05/2024

Client Sample No.: CCAL02 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PR068012.D Time Analyzed: 12:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.144	4.044	4.244	558.160	500.000	11.6
Aroclor-1016-2	4.163	4.063	4.263	565.080	500.000	13.0
Aroclor-1016-3	4.344	4.244	4.444	567.590	500.000	13.5
Aroclor-1016-4	4.396	4.296	4.496	559.810	500.000	12.0
Aroclor-1016-5	4.615	4.516	4.716	567.580	500.000	13.5
Aroclor-1260-1	5.721	5.621	5.821	570.010	500.000	14.0
Aroclor-1260-2	5.927	5.827	6.027	578.730	500.000	15.7
Aroclor-1260-3	6.087	5.988	6.188	556.530	500.000	11.3
Aroclor-1260-4	6.598	6.499	6.699	583.890	500.000	16.8
Aroclor-1260-5	6.864	6.764	6.964	550.600	500.000	10.1
Decachlorobiphenyl	8.333	8.235	8.435	55.830	50.000	11.7
Tetrachloro-m-xylene	3.011	2.911	3.111	55.400	50.000	10.8