

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2024 08/08/2024

Client Sample No.: CCAL03 Date Analyzed: 08/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR068106.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.902	4.800	5.000	516.290	500.000	3.3
Aroclor-1016-2	4.924	4.822	5.022	517.310	500.000	3.5
Aroclor-1016-3	4.989	4.887	5.087	526.210	500.000	5.2
Aroclor-1016-4	5.091	4.989	5.189	537.700	500.000	7.5
Aroclor-1016-5	5.406	5.303	5.503	534.520	500.000	6.9
Aroclor-1260-1	6.619	6.516	6.716	512.910	500.000	2.6
Aroclor-1260-2	6.901	6.798	6.998	510.370	500.000	2.1
Aroclor-1260-3	7.291	7.188	7.388	510.760	500.000	2.2
Aroclor-1260-4	7.533	7.430	7.630	516.580	500.000	3.3
Aroclor-1260-5	7.872	7.768	7.968	521.770	500.000	4.4
Decachlorobiphenyl	9.587	9.481	9.681	52.090	50.000	4.2
Tetrachloro-m-xylene	3.678	3.577	3.777	53.100	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2024 08/08/2024

Client Sample No.: CCAL03 Date Analyzed: 08/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR068106.D Time Analyzed: 09:08

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.139	4.041	4.241	529.420	500.000	5.9
Aroclor-1016-2	4.159	4.060	4.260	536.180	500.000	7.2
Aroclor-1016-3	4.339	4.241	4.441	565.530	500.000	13.1
Aroclor-1016-4	4.391	4.293	4.493	541.890	500.000	8.4
Aroclor-1016-5	4.611	4.513	4.713	540.600	500.000	8.1
Aroclor-1260-1	5.715	5.617	5.817	522.890	500.000	4.6
Aroclor-1260-2	5.922	5.823	6.023	534.720	500.000	6.9
Aroclor-1260-3	6.082	5.984	6.184	534.100	500.000	6.8
Aroclor-1260-4	6.594	6.494	6.694	529.970	500.000	6.0
Aroclor-1260-5	6.859	6.761	6.961	541.240	500.000	8.2
Decachlorobiphenyl	8.331	8.230	8.430	54.930	50.000	9.9
Tetrachloro-m-xylene	3.009	2.910	3.110	53.630	50.000	7.3