

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2024 08/08/2024

Client Sample No.: CCAL04 Date Analyzed: 08/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR068114.D Time Analyzed: 11:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.900	4.800	5.000	514.210	500.000	2.8
Aroclor-1016-2	4.922	4.822	5.022	519.770	500.000	4.0
Aroclor-1016-3	4.987	4.887	5.087	520.380	500.000	4.1
Aroclor-1016-4	5.089	4.989	5.189	525.830	500.000	5.2
Aroclor-1016-5	5.403	5.303	5.503	546.590	500.000	9.3
Aroclor-1260-1	6.616	6.516	6.716	519.900	500.000	4.0
Aroclor-1260-2	6.899	6.798	6.998	520.480	500.000	4.1
Aroclor-1260-3	7.288	7.188	7.388	518.560	500.000	3.7
Aroclor-1260-4	7.530	7.430	7.630	524.060	500.000	4.8
Aroclor-1260-5	7.868	7.768	7.968	526.760	500.000	5.4
Decachlorobiphenyl	9.583	9.481	9.681	52.580	50.000	5.2
Tetrachloro-m-xylene	3.677	3.577	3.777	52.890	50.000	5.8

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 08/08/2024 08/08/2024

Client Sample No.: CCAL04 Date Analyzed: 08/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR068114.D Time Analyzed: 11:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.141	4.041	4.241	487.670	500.000	-2.5
Aroclor-1016-2	4.159	4.060	4.260	494.210	500.000	-1.2
Aroclor-1016-3	4.340	4.241	4.441	519.380	500.000	3.9
Aroclor-1016-4	4.391	4.293	4.493	517.610	500.000	3.5
Aroclor-1016-5	4.611	4.513	4.713	520.450	500.000	4.1
Aroclor-1260-1	5.716	5.617	5.817	497.160	500.000	-0.6
Aroclor-1260-2	5.922	5.823	6.023	506.110	500.000	1.2
Aroclor-1260-3	6.082	5.984	6.184	506.420	500.000	1.3
Aroclor-1260-4	6.594	6.494	6.694	496.150	500.000	-0.8
Aroclor-1260-5	6.860	6.761	6.961	476.800	500.000	-4.6
Decachlorobiphenyl	8.330	8.230	8.430	42.120	50.000	-15.8
Tetrachloro-m-xylene	3.009	2.910	3.110	46.900	50.000	-6.2