

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD_D Date(s) Analyzed: 07/15/2024 07/15/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.88	5.78	5.98	6796630
		2	6.68	6.58	6.78	13778900
		3	7.10	7.00	7.20	57571800
		4	7.19	7.09	7.29	43296400
		5	7.61	7.51	7.71	29763900

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD_D Date(s) Analyzed: 07/15/2024 07/15/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	4.99	4.89	5.09	45686800
		2	5.67	5.57	5.77	46047900
		3	6.71	6.61	6.81	211144000
		4	7.03	6.93	7.13	78891900
		5	7.16	7.06	7.26	58828700