

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD_L Date(s) Analyzed: 07/15/2024 07/15/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.86	5.76	5.96	9513430
		2	6.45	6.35	6.55	12807300
		3	7.06	6.96	7.16	59009200
		4	7.15	7.05	7.25	45061100
		5	7.94	7.84	8.04	32101000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD_L Date(s) Analyzed: 07/15/2024 07/15/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.35	5.25	5.45	11276200
		2	5.70	5.60	5.80	11719400
		3	6.62	6.52	6.72	37283100
		4	6.75	6.65	6.85	54876100
		5	7.19	7.09	7.29	15293900