

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/05/2024 Initial Calibration Date(s): 08/05/2024 08/05/2024

Continuing Calib Time: 19:44 Initial Calibration Time(s): 15:08 16:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.05	9.06	8.96	9.16	0.01
Tetrachloro-m-xylene	3.54	3.55	3.45	3.65	0.01
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.77	4.78	4.68	4.88	0.01
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.08	5.98	6.18	0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.19	6.20	6.10	6.30	0.01
Endrin	6.58	6.58	6.48	6.68	0.00
Endosulfan II	6.79	6.80	6.70	6.90	0.01
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.02	7.03	6.93	7.13	0.01
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.65	7.55	7.75	0.01
Endrin aldehyde	6.92	6.93	6.83	7.03	0.01
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.95	5.85	6.05	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Continuing Calib Date: 08/05/2024 Initial Calibration Date(s): 08/05/2024 08/05/2024

Continuing Calib Time: 19:44 Initial Calibration Time(s): 15:08 16:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.93	7.93	7.83	8.03	0.00
Tetrachloro-m-xylene	2.79	2.79	2.69	2.89	0.00
alpha-BHC	3.29	3.29	3.19	3.39	0.00
beta-BHC	3.92	3.92	3.82	4.02	0.00
delta-BHC	4.15	4.15	4.05	4.25	0.00
gamma-BHC (Lindane)	3.62	3.62	3.52	3.72	0.00
Heptachlor	3.96	3.96	3.86	4.06	0.00
Aldrin	4.24	4.24	4.14	4.34	0.00
Heptachlor epoxide	4.74	4.74	4.64	4.84	0.00
Endosulfan I	5.11	5.11	5.01	5.21	0.00
Dieldrin	5.38	5.38	5.28	5.48	0.00
4,4'-DDE	5.25	5.25	5.15	5.35	0.01
Endrin	5.65	5.65	5.55	5.75	0.00
Endosulfan II	5.94	5.95	5.85	6.05	0.01
4,4'-DDD	5.80	5.80	5.70	5.90	0.00
Endosulfan sulfate	6.35	6.35	6.25	6.45	0.00
4,4'-DDT	6.05	6.05	5.95	6.15	0.00
Methoxychlor	6.62	6.63	6.53	6.73	0.01
Endrin ketone	6.85	6.86	6.76	6.96	0.01
Endrin aldehyde	6.12	6.13	6.03	6.23	0.01
alpha-Chlordane	5.06	5.06	4.96	5.16	0.01
gamma-Chlordane	4.99	5.00	4.90	5.10	0.01