

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG No.: P3390

Calibration Sequence : FE072924		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	225696676	132763	FE049298.D
850	112875714	132795	FE049299.D
340	39030081	114794	FE049300.D
170	23199215	136466	FE049301.D
85	12177347	143263	FE049302.D
AVG RF : 132016		% RSD : 7.981	AVG RT : 14.9946