

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG No.: P3390

Calibration Sequence : FF072324		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	232296862	136645	FF014300.D
850	95149660	111941	FF014301.D
340	33052807	97214	FF014302.D
170	19669849	115705	FF014303.D
85	10348820	121751	FF014304.D
AVG RF : 116651		% RSD : 12.324	AVG RT : 14.024