

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG No.: P3390
DataFile: FE049408.D Analyst Name: YP\AJ Analyst Date: 08-02-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
850	111535308	131218	132016	0.604