

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG No.: P3390
DataFile: FG013994.D Analyst Name: YP\AJ Analyst Date: 07-31-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
850	101188525	119045	117920	0.954