

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD\_R Calibration Date(s): 07/30/2024 07/30/2024  
Calibration Times: 10:41 11:19

GC Column: ZB-MR1 ID: 0.32 (mm)

| LAB FILE ID:               |             | CF 0.5 = <u>PR067871.D</u>  |             | CF 0.25 = <u>PR067872.D</u>  |             |             |          |
|----------------------------|-------------|-----------------------------|-------------|------------------------------|-------------|-------------|----------|
| CF 0.1 = <u>PR067873.D</u> |             | CF 0.05 = <u>PR067874.D</u> |             | CF 0.025 = <u>PR067875.D</u> |             |             |          |
| COMPOUND                   | CF 0.5      | CF 0.25                     | CF 0.1      | CF 0.05                      | CF 0.025    | CF          | %<br>RSD |
| DBCP                       | 88652500000 | 88233100000                 | 96053600000 | 99324100000                  | 88165000000 | 92085700000 | 6        |
| EDB                        | 44337800000 | 42908600000                 | 40888400000 | 39463000000                  | 44348000000 | 42389200000 | 5        |

**CALIBRATION FACTOR OF INITIAL CALIBRATION**

Contract: CHEM02

Lab Code: CHEM Case No.: P3390 SAS No.: P3390 SDG NO.: P3390

Instrument ID: ECD\_R Calibration Date(s): 07/30/2024 07/30/2024  
Calibration Times: 10:41 11:19

GC Column: ZB-MR2 ID: 0.32 (mm)

| LAB FILE ID:               |             | CF 0.5 = <u>PR067871.D</u>  |             | CF 0.25 = <u>PR067872.D</u>  |             |             |          |
|----------------------------|-------------|-----------------------------|-------------|------------------------------|-------------|-------------|----------|
| CF 0.1 = <u>PR067873.D</u> |             | CF 0.05 = <u>PR067874.D</u> |             | CF 0.025 = <u>PR067875.D</u> |             |             |          |
| COMPOUND                   | CF 0.5      | CF 0.25                     | CF 0.1      | CF 0.05                      | CF 0.025    | CF          | %<br>RSD |
| DBCP                       | 83776500000 | 85967100000                 | 93288400000 | 99175700000                  | 89275500000 | 90296600000 | 7        |
| EDB                        | 48274600000 | 47807000000                 | 48076400000 | 53184100000                  | 43555800000 | 48179600000 | 7        |